

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

JUN 22 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N		130195
Supervisor/Operator Performing the Verification Procedure:		
A	Name	Robert J. Ely
	ID#	2681
	Date	06-18-2009
	Agency	Craig Police Department
	Phone #	(907) 826-3330
	Instrument Location	506 2 nd Street, Craig Alaska 99921
Alco S/N		X172978
Target Value		.079
High Pressure		1200 PSI
Alco Test Value Average		.083
1 st Alco		.083
2 nd Alco		
Signature		Robert J. Ely
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 26th day of June 2009.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

- C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

JUNE 10, 2009

OPERATOR'S NAME:

ELY, RJ

OPERATOR'S NUMBER: 2681

SUBJECT'S LAST NAME:

USER OF CAL NEW ALCO

SUBJECT'S FIRST NAME/MI :

USER OF CAL

O.L. #: 123456789

DEPT/AGENCY: CRG1

CASE/REPORT: 2009-XXXX

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: X172978

BREATH ANALYSIS

.079 ADJUSTED FOR 30.14 in

ALCO TARGET .079 12:11

BLANK TEST .000 12:12

INTERNAL STANDARD VERIFIED 12:12

ALCO TV 30.14 in .083 12:13

BLANK TEST .000 12:14

SUBJECT SAMPLE .000 12:14

BLANK TEST .000 12:15

ALCO TV 30.14 in .083 12:15

BLANK TEST .000 12:16

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

JUNE 10, 2009

TIME 12:17

DIAGNOSTIC CHECK

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.14 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ;
@ A B C D E F G H I J K L M N O P Q R S T U V W X Y Z \] ^ _ ` a b c d e f g h i j k l m n o p q r s t u v w x y z { | } ~